



## *Accredited Laboratory*

A2LA has accredited

### **CERTIFIED SCALE INC.**

*Menomonee Falls, WI*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 30<sup>th</sup> day of November 2016.

A handwritten signature in black ink, written over a horizontal line.

President and CEO  
For the Accreditation Council  
Certificate Number 2006.01  
Valid to December 31, 2018

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*



SCOPE OF ACCREDITATION TO ISO 17025:2005

CERTIFIED SCALE INC.  
 N57 W13640 Carmen Avenue  
 Menomonee Falls, WI 53051  
 Jason Syverud Phone: 262 781 9290

CALIBRATION

Valid To: December 31, 2018

Certificate Number: 2006.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Mechanical

| Parameter/Equipment                         | Range            | CMC <sup>2</sup> (±)   | Comments  |                             |
|---|------------------|--|---|-----------------------------|
| Balances <sup>3</sup> – Class I             | (0 to 200) g     | 0.000095 g   | NIST Handbook 44 with ASTM E617 Class 0 and 1 standards |                             |
|   | (>200 to 400) g  | 0.0012 g   |   |                             |
|   | (>400 to 600) g  | 0.0022 g   |   |                             |
|   | (>600 to 800) g  | 0.0027 g   |   |                             |
|   | (>800 to 1000) g | 0.0032 g   |   |                             |
|   | (>1 to 10) kg    | (>1 to 10) kg  | 0.15 g  | ASTM E617 Class 2 standards |
|   |                  | (>10 to 30) kg   | 0.27 g  |                             |
|   |                  | (>30 to 60) kg   | 0.38 g  |                             |
|   |                  | (>60 to 100) kg  | 0.87 g  |                             |
|   |                  | (>100 to 200) kg   | 1.8 g   |                             |
| (>200 to 300) kg                            | 14 g             | ASTM E617 Class 2 standards and NIST H-105-1 Class F standards |   |                             |
| Scales and Balances <sup>3</sup> – Class II | (0 to 200) g     | 0.0027 g   | NIST Handbook 44 with                                   |                             |
|   | (>200 to 400) g  | 0.0066 g   |   |                             |
|   | (>400 to 600) g  | 0.013 g  | ASTM E617 Class 1 standards                             |                             |
|   | (>600 to 800) g  | 0.014 g  |   |                             |
|   | (>800 to 1000) g | 0.014 g  |   |                             |
|   |                  |  |   |                             |

| Parameter/Equipment  | Range   | CMC <sup>2</sup> (±)  | Comments  |
|--|---|---|---|
| Scales and Balances <sup>3</sup> –<br><br>Class II (cont)  | (>1 to 10) kg<br>(>10 to 30) kg<br>(>30 to 100) kg<br>(>100 to 200) kg<br><br>(>200 to 300) kg  | 0.15 g<br>0.67 g<br>1.4 g<br>2.9 g<br><br>15 g  | ASTM E617 Class 2 standards<br><br>ASTM E617 Class 2 standards and NIST H-105-1 Class F standards   |
| Scales <sup>3</sup> –<br><br>Class III<br><br><br><br><br><br><br><br><br><br><br><br><br><br><br><br><br><br><br>Class IIIL | (0 to 5) lb<br>(>5 to 10) lb<br>(>10 to 20) lb<br>(>20 to 50) lb<br>(>50 to 100) lb<br>(>100 to 200) lb<br>(>200 to 400) lb<br>(>400 to 1000) lb<br>(>1000 to 5000) lb<br>(>5000 to 10 000) lb<br>(>10 000 to 20 000) lb<br><br>(>20 000 to 50 000) lb<br>(>50 000 to 100 000) lb<br>(>100 000 to 200 000) lb | 0.00088 lb<br>0.0018 lb<br>0.0035 lb<br>0.0088 lb<br>0.018 lb<br>0.035 lb<br>0.08 lb<br>0.18 lb<br>0.88 lb<br>1.8 lb<br>3.5 lb<br><br>27 lb<br>37 lb<br>74 lb | NIST Handbook 44 with NIST H-105-1 Class F standards<br><br><br><br><br><br><br><br><br><br><br><br><br><br><br><br>NIST Handbook 44 with NIST H-105-1 Class F standards & substitution load(s) |

<sup>1</sup> This laboratory offers commercial calibration services at its facilities or at the client's site.

<sup>2</sup> Calibration and Measurement Capability uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup>Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

A handwritten signature in blue ink, appearing to be 'L. S. L.', is located at the bottom center of the page.